


2025 年 2 月 永瀬ダム 自動濁度装置 測定結果一覧表


※濁度測定器は、ホルマジン標準液（FTU）を使用して校正しています。

自動濁度計設置場所

	1 日		2 日		3 日		4 日		5 日		6 日		7 日		8 日		9 日		10 日		11 日		12 日		13 日		14 日		15 日			
水深	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度		
0.5	7.4	1.0	7.5	0.8	7.8	0.8	7.5	0.8	7.3	0.8	7.2	1.0	7.1	1.0	7.1	1.3	7.0	1.3	7.0	1.3	7.0	1.6	6.9	1.3	6.8	1.3	6.8	1.3	6.9	1.0		
1	7.4	1.0	7.5	0.8	7.8	0.8	7.5	0.8	7.3	0.8	7.2	1.0	7.1	1.0	7.1	1.3	7.0	1.3	7.0	1.6	7.0	1.3	6.9	1.6	6.8	1.3	6.8	1.3	6.9	1.0		
2	7.4	0.8	7.5	0.8	7.8	1.0	7.5	0.8	7.3	0.8	7.2	1.3	7.1	1.0	7.1	1.3	7.0	1.3	6.9	1.3	7.0	1.6	6.9	1.3	6.8	1.3	6.8	1.3	6.9	1.3		
3	7.4	0.8	7.5	1.0	7.8	1.0	7.5	0.8	7.2	0.8	7.2	1.6	7.1	1.0	7.1	1.3	7.0	1.6	6.9	1.3	6.9	1.6	6.9	1.3	6.8	1.3	6.8	1.3	6.9	1.0		
4	7.4	1.0	7.5	0.8	7.8	1.0	7.5	0.8	7.2	1.0	7.2	1.3	7.1	1.0	7.1	1.3	7.0	1.6	7.0	1.3	6.9	1.3	6.9	1.6	6.8	1.3	6.8	1.6	6.9	1.2		
5	7.4	0.8	7.5	1.0	7.8	1.0	7.5	1.0	7.2	0.8	7.2	1.2	7.1	1.2	7.1	1.3	7.0	1.3	7.0	1.3	6.9	1.3	6.9	1.3	6.9	1.3	6.8	1.3	6.9	1.3		
6	7.4	1.2	7.5	0.8	7.7	1.0	7.5	1.0	7.2	0.8	7.1	1.3	7.1	1.0	7.0	1.6	7.0	1.6	7.0	1.6	7.0	1.8	6.9	1.3	6.9	1.3	6.8	1.3	6.9	1.3		
7	7.4	1.2	7.4	0.8	7.5	1.0	7.4	1.0	7.2	0.8	7.2	1.3	7.1	1.0	7.0	1.3	7.0	1.3	7.0	1.3	7.0	1.6	6.9	1.6	6.9	1.3	6.8	1.3	6.9	1.3		
8	7.4	1.0	7.4	1.0	7.5	1.0	7.4	1.0	7.2	1.0	7.1	1.3	7.1	1.0	7.0	1.3	7.0	1.6	7.0	1.6	7.0	1.6	6.9	1.6	6.9	1.3	6.8	1.3	6.9	1.3		
9	7.3	1.0	7.4	1.0	7.4	1.0	7.4	1.3	7.3	1.0	7.1	1.3	7.1	1.0	7.0	1.3	7.0	1.6	7.0	1.6	7.0	1.6	6.9	1.8	6.9	1.3	6.8	1.3	6.9	1.3		
10	7.3	1.3	7.3	1.0	7.4	0.8	7.3	1.3	7.2	1.3	7.1	1.3	7.1	1.3	7.0	1.6	7.0	1.6	7.0	1.6	7.0	1.6	6.9	1.6	6.9	1.3	6.8	1.3	6.9	1.6		
12	7.3	1.0	7.3	1.3	7.3	2.2	7.3	2.7	7.2	1.3	7.1	2.0	7.1	1.0	7.0	1.3	7.0	1.6	7.0	2.4	6.9	2.2	6.9	1.6	6.9	1.3	6.8	1.6	6.8	1.6		
14	7.3	1.0	7.2	2.2	7.3	3.0	7.2	2.7	7.2	2.4	7.1	3.0	7.1	2.0	7.0	1.6	7.0	2.0	7.0	3.0	6.9	2.0	6.9	1.6	6.8	1.3	6.8	1.3	6.8	1.6		
16	7.2	1.6	7.2	4.5	7.2	4.7	7.2	3.5	7.1	2.4	7.1	3.7	7.1	2.0	7.0	2.0	7.0	2.4	7.0	3.9	6.9	2.0	6.9	1.8	6.8	1.6	6.8	1.3	6.8	2.0		
18	7.2	2.4	7.1	5.8	7.2	5.1	7.2	3.9	7.1	3.5	7.1	4.1	7.1	2.7	7.0	2.7	7.0	3.0	7.0	3.9	6.9	2.0	6.9	2.0	6.8	2.4	6.8	1.3	6.8	2.2		
20	7.2	4.1	7.1	5.4	7.1	5.1	7.1	5.1	7.1	4.1	7.1	3.7	7.1	3.2	7.0	3.2	7.0	4.5	6.9	4.3	6.9	2.4	6.9	2.0	6.8	3.0	6.8	1.6	6.8	2.0		
25	7.1	6.4	7.1	5.6	7.1	5.6	7.1	4.5	7.1	4.5	7.0	3.5	7.0	3.0	7.0	4.9	6.9	4.5	6.9	3.5	6.9	3.2	6.8	3.7	6.7	4.7	6.8	2.0	6.7	4.5		
30	7.1	7.8	7.0	4.1	7.0	4.7	7.0	4.5	7.0	4.5	7.0	3.9	7.0	5.6	6.9	7.8	6.9	5.6	6.8	4.3	6.9	4.5	6.8	3.5	6.6	4.1	6.7	3.7	6.6	5.4		
35	7.1	7.8	7.0	5.6	7.0	5.6	6.9	6.1	7.0	6.1	7.0	6.1	7.0	8.9	6.9	7.6	6.8	5.8	6.8	6.6	6.8	7.5	6.7	6.0	6.6	5.1	6.6	5.8	6.6	5.6		
40	7.0	7.8	6.9	10.7	6.9	8.3	6.9	8.9	6.9	8.3	6.9	13.3	7.0	12.9	6.9	8.0	6.8	8.9	6.7	12.4	6.7	11.2	6.7	6.1	6.6	7.3	6.6	12.1	6.6	8.5		
	16 日		17 日		18 日		19 日		20 日		21 日		22 日		23 日		24 日		25 日		26 日		27 日		28 日		29 日		30 日		31 日	
水深	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度	水温	濁度		
0.5	7.3	0.8	7.3	0.8	7.3	0.5	7.1	0.8	7.0	0.8	6.9	0.3	6.9	0.3	6.8	0.3	6.8	0.3	6.6	0.5	6.9	1.8	7.0	0.3	7.3	0.3						
1	7.3	0.8	7.3	0.8	7.2	0.8	7.1	0.3	7.0	0.3	6.9	0.3	6.9	0.3	6.8	0.3	6.8	0.3	6.6	0.3	6.9	0.3	7.0	0.3	7.3	0.3						
2	7.3	0.8	7.3	0.8	7.2	0.8	7.1	0.8	7.0	0.8	6.9	0.3	6.9	0.8	6.8	0.5	6.8	0.3	6.6	0.3	6.9	0.3	7.0	0.3	7.3	0.3						
3	7.3	1.0	7.3	0.8	7.2	0.8	7.1	0.8	7.0	0.3	6.9	0.5	6.9	0.3	6.8	0.5	6.8	0.3	6.6	0.8	6.9	0.3	7.0	0.3	7.3	0.5						
4	7.3	1.0	7.3	0.8	7.2	0.8	7.1	0.5	7.0	1.3	6.9	0.8	6.9	0.8	6.8	0.8	6.8	0.3	6.6	0.8	6.9	0.3	7.0	0.3	7.3	0.3						
5	7.3	1.0	7.3	0.8	7.2	0.8	7.1	0.3	7.0	0.3	6.9	0.8	6.9	0.8	6.8	0.8	6.8	0.3	6.6	0.8	6.9	0.8	7.0	0.8	7.3	0.3						
6	7.3	1.2	7.3	1.0	7.2	0.8	7.1	0.3	7.0	0.8	7.0	0.8	6.9	0.3	6.8	0.8	6.8	0.3	6.6	0.8	6.9	0.8	7.0	0.8	7.3	0.5						
7	7.1	1.2	7.1	1.0	7.2	1.0	7.1	0.8	7.0	0.8	7.0	0.8	6.9	0.8	6.8	0.8	6.8	0.3	6.6	0.8	6.9	0.8	7.0	0.8	7.2	0.5						
8	6.9	1.3	7.0	1.0	7.1	1.0	7.0	1.0	6.9	0.8	7.0	0.8	6.9	0.5	6.8	0.8	6.8	0.8	6.6	0.8	6.8	0.8	7.0	0.8	7.2	0.5						
9	6.9	1.3	7.0	1.2	7.1	1.0	6.9	0.8	6.9	0.8	7.0	0.8	6.9	0.8	6.8	0.5	6.8	0.3	6.6	0.5	6.7	0.8	6.9	0.8	7.1	0.8						
10	6.9	1.3	6.9	1.3	7.0	1.0	6.8	0.8	6.8	0.8	6.9	0.8	6.9	0.8	6.8	0.8	6.8	0.8	6.6	0.8	6.7	0.8	6.8	0.8	6.9	0.8						
12	6.8	1.6	6.8	1.3	6.8	1.3	6.8	1.0	6.8	1.0	6.9	0.8	6.8	0.8	6.8	0.5	6.7	1.0	6.6	0.8	6.6	0.8	6.8	0.8	6.8	1.0						
14	6.8	1.8	6.8	1.3	6.8	1.3	6.7	1.2	6.7	1.0	6.8	0.8	6.8	0.8	6.7	1.2	6.7	1.2	6.6	0.8	6.6	1.0	6.7	1.3	6.7	1.3						
16	6.8	1.8	6.7	2.0	6.7	1.3	6.7	1.3	6.7	1.2	6.7	1.3	6.7	0.8	6.7	1.3	6.6	1.3	6.6	1.3	6.6	1.3	6.6	2.2	6.6	1.6						
18	6.7	3.0	6.7	2.9	6.7	1.6	6.7	1.6	6.7	1.8	6.7	1.8	6.7	1.3	6.7	1.3	6.6	1.3	6.6	1.6	6.6	2.4	6.6	3.0	6.6	2.0						
20	6.7	3.9	6.7	2.7	6.7	1.8	6.6	2.2	6.7	2.0	6.7	1.8	6.7	1.3	6.6	1.3	6.6	1.8	6.6	2.2	6.6	3.0	6.6	4.5	6.6	3.5						
25	6.6	5.1	6.6	3.0	6.7	3.5	6.6	2.7	6.6	2.4	6.6	2.2	6.6	2.0	6.6	2.0	6.6	3.5	6.6	3.5	6.6	5.6	6.6	4.5	6.6	3.9						
30	6.6	4.1	6.6	3.5	6.6	3.5	6.6	3.0	6.5	3.5	6.6	2.4	6.5	2.9	6.5	4.7	6.6	8.5	6.5	4.9	6.5	5.6	6.5	3.9	6.5	5.4						
35	6.6	5.4	6.5	5.6	6.5	7.0	6.5	6.8	6.5	5.1	6.5	8.3	6.6	7.6	6.6	11.0	6.6	10.7	6.5	8.5	6.5	7.6	6.5	9.9	6.5	6.8						
40	6.5	6.8	6.5	9.3	6.5	10.4	6.5	11.8	6.5	8.5	6.5	10.6	6.6	12.6	6.6	14.8	6.6	15.0	6.6	15.8	6.5	14.5	6.5	12.9	6.5	7.8						



50
45
40
35
30
25
20
15
10
5
(FTU)



▲永瀬ダム周辺 位置図



▲自動濁度計 全景

永瀬ダム堰堤上流付近に台船を浮かべ、その上に自動で濁度を取得する装置を設置しています。

毎日朝9時になると測定器を沈め、設定された深度における濁度を計測しています。

